

Hengst SE  
Nienkamp 55-85 · 48147 Münster  
Postfach 20 14 54 · 48095 Münster  
Germany

T +49 (0)251 20202-0  
F +49 (0)251 20202-900

info@hengst.de  
www.hengst.de

## Product testing MNS 03 according to DIN EN 14683

According to the investigations by Hohenstein Laboratories GmbH & Co. KG we confirm that you have passed the following tests:

- Bacterial filtration efficiency:  $\geq 98\%$
- Differential pressure:  $< 40 \text{ Pa/cm}^2$
- Microbial cleanliness:  $\leq 30 \text{ KBE/g}$


The test for biocompatibility shows that no cytotoxic substances are released.

The tested probe does fulfil the requirements of DIN EN 14683 for medical masks type II.

Münster, 06/15/2020



Christopher Heine  
CEO



ppa. Oliver Nabrotzky  
Group Vice President